


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0100</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>Johnson Matthey plc - Noble Metals</p> <p>Issue No: 016 Issue date: 21 July 2010</p>	
	<p>Orchard Road Royston Herts SG8 5HE</p>	<p>Contact: Nick Mahoney Tel: +44(0) 1763 253893 Fax: +44 (0)1763 253451 E-Mail: nick.mahoney@matthey.com Website: www.noble.matthey.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE			
Platinum thermocouples			
Fixed point measurements	0 °C to 1100 °C 1100 °C to 1600 °C 1600 °C to 1800 °C	1.0 °C 2.0 °C 3.0 °C	
Measurement by comparison	0 °C to 1100 °C 1100 °C to 1400 °C	2.0 °C 3.0 °C	
<p>END</p>			